

EC Certificate of Conformity

The Notified Body

MEDCERT Zertifizierungs- und Prüfungsgesellschaft für die Medizin GmbH
Pilatuspool 2 – 20355 Hamburg – Germany

herewith certifies that the company:

MADESPA, S.A.
Calle Rio Jarama, 120
45007 Toledo
Spain

has introduced, applies and maintains a quality assurance system for the products / product categories listed in the appendix.

The compliance of this quality assurance system with the below mentioned requirements of the **Council Directive 93/42/EEC** was verified by an audit:

Annex V

This certification is subject to surveillance by MEDCERT.

Effective date: 2020-10-23

Expiry date: 2023-08-29

Report No.: 7487PS02F

Process No.: QS – 7487

Certificate No.: 7487GB414201023A

Hamburg, 2020-10-23



MEDCERT Certification Body
(Dr. Andreas Schich)

The certificate is only valid when provided entirely with all of its pages.
To verify the validity of this certificate, contact info@medcert.de.

MEDCERT Identification Number: 0482

Form F10010005e EN / Rev. 11 / 2019.11.14



Benannt durch/Designated by
Zentralsstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-237.10.15

Appendix of EC Certificate of Conformity

Process No.: QS – 7487

Certificate No.: 7487GB414201023A

List of products / product categories included in the scope of certificate

Dental Materials:

- Core build-up materials
- Dentine and enamel adhesives
- Luting materials (polymers)
- Materials for fixed prosthesis (inlays, onlays, veneers) (polymers)
- Materials for surface preparation (etch)
- Plastic materials for direct insertion (polymers)

– End of list –

This appendix is integral part of the above-referenced certificate.
The certificate is only valid when provided entirely with all of its pages.
To verify the validity of this certificate, contact info@medcert.de.

MEDCERT Identification Number: 0482

Form F10010005e EN / Rev. 11 / 2019.11.14



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-237.10.15

	Manufacturer's Self-Declaration	RV:00
		Fecha: 12/07/2023
		Page 1 de 4

In relation to Regulation 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices, in particular with respect to

- the validity of certificates issued under Council Directive 93/42/EEC on Medical Devices (MDD) (Directive Certificates) *and/or*¹
- the compliance of the devices and us as their manufacturer with the conditions for the continued placing on the market and putting into service

Manufacturer name	MADESPA S.A.
Manufacturer address and contact details	C/RÍO JARAMA 120, CP 45007 TOLEDO – TOLEDO SPAIN
Single Registration Number (SRN) (if available)	ESMF-000003915

Authorised Representative name (if applicable)	NA
Authorised Representative address and contact details	NA
Single Registration Number (SRN) (if available)	NA

Notified body name (if applicable)	MEDCERT
Notified body number (if applicable)	0482
Directive Certificate number(s) to which this confirmation is made (if applicable)	7487GB414201023A
Original expiry date as indicated on the Directive Certificate prior to the extension of the validity (if applicable)	29/08/2023
End date of extended validity/transition period	31/12/2028

¹ The first condition is not applicable in case of devices for which the conformity assessment procedure pursuant to MDD did not require the involvement of a notified body, for which the declaration of conformity was drawn up prior to 26 May 2021 and for which the conformity assessment procedure pursuant to this Regulation requires the involvement of a notified body.

	Manufacturer's Self-Declaration	RV:00
		Fecha: 12/07/2023
		Page 2 de 4

We, as the manufacturer declare under our sole responsibility:

- for the above listed **Directive Certificate** (or see attached schedule, if multiple certificates) the conditions for the legal extension of validity as required in Article 120.2 of the MDR are met *and/or*²
- the listed **device(s)** in the attached schedule and we as their manufacturer are in compliance with the conditions listed in Article 120.3c of the MDR for continued placing on the market and putting into service,

namely by fulfilling the following conditions:

➤ **Directive Certificate(s)** as listed above or in the attached schedule

- Directive Certificate(s) covering the listed device(s) was/were issued after 25 May 2017, was/were valid on 26 May 2021, was/were not withdrawn by 20 March 2023

- *Choose applicable statements:*


☐ Expired *before* 20 March 2023:

- ☐ Before the original date of expiry as indicated on the Directive Certificate, we and the notified body have signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII to this Regulation for the conformity assessment in respect of the device covered by the expired certificate or in respect of a device intended to substitute that device
- ☐ A Competent Authority has granted a derogation from the applicable conformity assessment procedure in accordance with Article 59(1) MDR (may be provided upon request)
- ☐ A Competent Authority has required the manufacturer, in accordance with Article 97(1) MDR, to carry out the applicable conformity assessment procedure (may be provided upon request)

☐ Expired/expires *after* 20 March 2023:

- ☒ A formal application to the notified body in accordance with Section 4.3, first subparagraph of Annex VII MDR for conformity assessment has been made or will be made/submitted by us to a notified body no later than 26 May 2024 for the device(s) listed in the attached schedule or its substitute and a signed written agreement is/will be in place in accordance with Section 4.3, second subparagraph of Annex VII MDR before 26 September 2024.
- ☐ We do not intent to lodge an application for conformity assessment by 26 May 2024, therefore the transition period will end on 26 May 2024.

² The first condition is not applicable in case of devices for which the conformity assessment procedure pursuant to MDD did not require the involvement of a notified body, for which the declaration of conformity was drawn up prior to 26 May 2021 and for which the conformity assessment procedure pursuant to this Regulation requires the involvement of a notified body

	Manufacturer's Self-Declaration	RV:00
		Fecha: 12/07/2023
		Page 3 de 4

➤ **Upclassified devices**

In case of devices for which the conformity assessment procedure pursuant to MDD did not require the involvement of a notified body, for which the declaration of conformity was drawn up prior to 26 May 2021 and for which the conformity assessment procedure pursuant to this Regulation requires the involvement of a notified body:

Choose one applicable statement:

- ☐ A formal application to the notified body in accordance with Section 4.3, first subparagraph of Annex VII MDR for conformity assessment has been made or will be made/submitted by us to a notified body no later than 26 May 2024 for the device(s) listed in the attached schedule or its substitute and a signed written agreement is/will be in place in accordance with Section 4.3, second subparagraph of Annex VII MDR before 26 September 2024.
- ☐ We do not intent to lodge an application for conformity assessment by 26 May 2024, therefore the transition period will end on 26 May 2024.

➤ **Quality Management System (QMS)**

• *Choose one applicable statement:*

- ☐ A QMS in accordance with Article 10(9) MDR will be put in place by no later than 26 May 2024.
- ☒ A QMS in accordance with Article 10(9) MDR is in place.
- ☐ A notified body has issued the attached certificate for the MDR-compliant QMS.

➤ **Device(s) as listed in the attached schedule**

- The device(s) continue to comply with the AIMDD or MDD.
- The device(s) has/have not been significantly changed in its/their design and intended purpose since 26 May 2021.
- The device(s) do not present an unacceptable risk to health or safety of patients, users or other persons, or to other aspects of the protection of public health.


Signed for and on behalf of the manufacturer:

Full Company Name MADESPA S.A.

Location & Date Toledo, 12/07/2023

Jacob Martín Ruano
General Manager


Jacob Martín Ruano
General Manager
 Tel. +34.925.241.025

	Manufacturer's Self-Declaration	RV:00
		Fecha: 12/07/2023
		Página 4 de 4

Schedule of Devices

The above Manufacturer's Declaration is valid for the following devices:

Identification of the device Device name / Basic UDI-DI (under MDR application)	MDR Device classification	If the MDR device is a substitute device, identification of the corresponding MDD device	MDD Certificate Reference(s) of the devices under MDR application, and the NB Identification (Name and number)	Original expiry date as indicated on the Directive Certificate prior to the extension of the validity (if applicable)	End date of extended validity/transition period
COMPOSITES BASIC UDI-DI: 84353161DM000105T	IIa	Core build-up materials Materials for fixed prosthesis (inlays, onlays, veneers) (polymers) Plastic materials for direct insertion (polymers)	Certificate N° 7487GB414201023A Issued by Medcert 0482	29/08/2023	31/12/2028
ADHESIVES BASIC UDI-DI: 84353161DM000205W	<u>IIa</u>	Dentine and enamel adhesives	Certificate N° 7487GB414201023A Issued by Medcert 0482	29/08/2023	31/12/2028
ETCHING GEL BASIC UDI-DI: 84353161DM000305Z	<u>IIa</u>	Materials for surface preparation (etch)	Certificate N° 7487GB414201023A Issued by Medcert 0482	29/08/2023	31/12/2028

- End of list -